

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT

ME0525A

2025-07-10

Instrument	Parameter	Low		Normal		High	
		LOT	ME0525A-1	LOT	ME0525A-2	LOT	ME0525A-3
element HT5+ QC Mode	Ret# ×10 ³ /μL	75.9	± 64.4	227.5	± 99.0	205.9	± 99.0
	Ret%	1.30	± 1.12	4.81	± 2.00	8.65	± 3.50
	IRF %	10.4	± 15.0	11.8	± 13.0	9.9	± 25.0
	LFR %	89.6	± 15.0	88.2	± 15.0	90.1	± 25.0
	MFR %	8.4	± 12.0	10.1	± 10.0	7.7	± 22.0
	HFR %	2.0	± 10.0	1.7	± 10.0	2.2	± 10.0
	RHE pg	26.5	± 12.0	22.7	± 12.0	19.8	± 12.0
	IPF %	7.8	± 15.0	7.7	± 15.0	7.5	± 15.0

Before using, refer to the instruction sheet for mixing directions.