

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOTME0725A

2025-09-10

Instrument	Parameter	Low		Normal		High	
		LOT	ME0725A-1	LOT	ME0725A-2	LOT	ME0725A-3
element HT5+ QC Mode	Ret# ×10 ³ /μL	82.5	± 64.4	235.9	± 99.0	191.3	± 99.0
	Ret%	1.42	± 1.12	5.04	± 2.00	8.28	± 3.50
	IRF %	9.0	± 15.0	13.1	± 13.0	9.3	± 25.0
	LFR %	91.0	± 15.0	86.9	± 15.0	90.7	± 25.0
	MFR %	6.0	± 12.0	11.5	± 10.0	6.6	± 22.0
	HFR %	3.0	± 10.0	1.6	± 10.0	2.7	± 10.0
	RHE pg	25.2	± 12.0	23.4	± 12.0	19.9	± 12.0
	IPF %	7.6	± 15.0	7.5	± 15.0	7.7	± 15.0

Before using, refer to the instruction sheet for mixing directions.