

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOTME1125A

2026-01-10

Instrument	Parameter	Low		Normal		High	
		LOT	ME1125A-1	LOT	ME1125A-2	LOT	ME1125A-3
element HT5+ QC Mode	Ret# ×10 ³ /μL	92.0	± 64.4	227.5	± 99.0	189.6	± 99.0
	Ret%	1.58	± 1.12	4.84	± 2.00	8.39	± 3.50
	IRF %	9.8	± 15.0	12.9	± 13.0	10.4	± 25.0
	LFR %	90.2	± 15.0	87.1	± 15.0	89.6	± 25.0
	MFR %	7.3	± 12.0	11.4	± 10.0	8.1	± 22.0
	HFR %	2.5	± 10.0	1.5	± 10.0	2.3	± 10.0
	RHE pg	24.9	± 12.0	22.4	± 12.0	19.2	± 12.0
	IPF %	7.6	± 15.0	7.7	± 15.0	7.6	± 15.0

Before using, refer to the instruction sheet for mixing directions.