

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT ME0526A



2026-07-10

Instrument	Parameter	Low		Normal		High	
		LOT	ME0526A-1	LOT	ME0526A-2	LOT	ME0526A-3
element HT5+	Ret# x10 ³ /μL	63.4	± 64.4	220.4	± 99.0	185.5	± 99.0
QC Mode	Ret%	1.08	± 1.12	4.67	± 2.00	8.32	± 3.50
	IRF %	11.3	± 15.0	11.2	± 13.0	7.7	± 25.0
	LFR %	88.7	± 15.0	88.8	± 15.0	92.3	± 25.0
	MFR %	9.4	± 12.0	9.3	± 10.0	4.4	± 22.0
	HFR %	1.9	± 10.0	1.9	± 10.0	3.3	± 10.0
	RHE pg	24.3	± 12.0	23.2	± 12.0	20.4	± 12.0
	IPF %	7.6	± 15.0	7.6	± 15.0	7.6	± 15.0

Before using, refer to the instruction sheet for mixing directions.