

BC-RET

HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT

ME0926A

2026-11-10

Instrument	Parameter	Low		Normal		High	
		LOT	ME0926A-1	LOT	ME0926A-2	LOT	ME0926A-3
element HT5+ QC Mode	Ret# ×10 ³ /μL	86.1	± 64.4	229.3	± 99.0	178.1	± 99.0
	Ret%	1.47	± 1.12	5.04	± 2.00	8.17	± 3.50
	IRF %	12.3	± 15.0	12.2	± 13.0	8.5	± 25.0
	LFR %	87.7	± 15.0	87.8	± 15.0	91.5	± 25.0
	MFR %	10.6	± 12.0	10.4	± 10.0	5.4	± 22.0
	HFR %	1.7	± 10.0	1.8	± 10.0	3.1	± 10.0
	RHE pg	24.5	± 12.0	22.9	± 12.0	21.2	± 12.0
	IPF %	7.6	± 15.0	7.6	± 15.0	7.5	± 15.0

Before using, refer to the instruction sheet for mixing directions.